

DEC 09 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130177✓
Name		STEVE KEVAN	ID# 6585✓
		Date	11/12/10✓
A	Agency	ALASKA STATE TROOPERS	Phone # 438-2029
Instrument Location		ST. MARY'S POST	
Alco S/N		X172959✓	Target Value .081✓
			High Pressure 0950
B	Alco Test Values	.081✓	.082✓
		1st Alco	2nd Alco
Signature		[Signature]	

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 03RD day of DEC, 2010.

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130177

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

NOVEMBER 12, 2010

OPERATOR'S NAME:

KEVIN / STEVE

OPERATOR'S NUMBER: 6585

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/NI :

TEST

O.L. #: 00

DEPT/AGENCY: SMV8

CASE/REPORT: 10-107507

TEST TYPE: U

ALCO TARGET VALUE: .001

ALCO S/N: X172959

— BREATH ANALYSIS —

.001 ADJUSTED FOR 29.38 in

ALCO TARGET	.079	15:32
BLANK TEST	.000	15:33
INTERNAL STANDARD	VERIFIED	15:33
ALCO TO 29.38 in	.001	15:34
BLANK TEST	.000	15:34
SUBJECT SAMPLE	.000	15:35
BLANK TEST	.000	15:35
ALCO TO 29.38 in	.002	15:36
BLANK TEST	.000	15:37

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

NOVEMBER 12, 2010

TIME 15:28

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 29.38 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~